



West End Grange
 Epworth Water Gardens
 West End Road
 Epworth
 Doncaster
 DN9 1LA



Certificate Number:

2018763

Tel: 01427 875 867
 Web: www.allweighs.co.uk

Calibration Certificate of Accuracy

Business Name: Attero

Address: Foster Street
Hull

Postcode: HU8 8BT

Contact Name: Martin

Test Date: 4 December 2018

Location: Weighbridge

Manufacturer: Dini Argeo

Serial/Plant No: 63306384

Capacity: 50000 kg

Increment: 20 kg

Frequency: 12 months

Next due: December 2019

As Found

Measurement Points	Reading	Deviation
0	0	0
8000	8000	0
16000	16000	0
23200	23180	20
33200	33180	20
43860	43800	60

Repeatability

43860	43800	60
-------	-------	----

Off-centre Loading

10000	9980	20
-------	------	----

Engineer Name: J Scott

Definitive

Measurement Points	Reading	Deviation
0	0	0
8000	8000	0
16000	16000	0
23200	23200	0
33200	33200	0
43860	44860	0

Repeatability

43860	43860	0
-------	-------	---

Off-centre Loading


10000	10000	0
-------	-------	---

Trading Standards Maximum Tolerance

60 +/- 60 kg

Traceable Weights Used: W18012

Calibration method and equipment used traceable to national standards.



West End Grange
 Epworth Water Gardens
 West End Road
 Epworth
 Doncaster
 DN9 1LA



Certificate Number:

2018726

Tel: 01427 875 867
 Web: www.allweighs.co.uk

Calibration Certificate of Accuracy

Business Name: Attero

Address: Gibson Lane
North Ferriby
Melton

Postcode: HU14 3HH

Contact Name: Martin

Test Date: 4 December 2018

Location: Out Weighbridge

Manufacturer: Dini Argeo

Serial/Plant No: 0100

Capacity: 50000 kg

Increment: 20 kg

Frequency: 12 months

Next due: December 2019

As Found

Measurement Points	Reading	Deviation
0	0	0
8000	8000	0
16000	16000	0
23200	23200	0
33200	33200	0
43860	43840	20

Repeatability

43860	43840	20
-------	-------	----

Off-centre Loading

10000	10000	0
-------	-------	---

Engineer Name: J Scott

Definitive

Measurement Points	Reading	Deviation
0	0	0
8000	8000	0
16000	16000	0
23200	23200	0
33200	33200	0
43860	43860	0

Repeatability

43860	43860	0
-------	-------	---

Off-centre Loading

10000	10000	0
-------	-------	---

Trading Standards Maximum Tolerance

60	+/-	60	kg
----	-----	----	----

Traceable Weights Used:

W18012

Calibration method and equipment used traceable to national standards.



West End Grange
 Epworth Water Gardens
 West End Road
 Epworth
 Doncaster
 DN9 1LA



Certificate Number:

2018727

Tel: 01427 875 867
 Web: www.allweighs.co.uk

Calibration Certificate of Accuracy

Business Name: Attero

Address: Gibson Lane
North Ferriby
Melton

Postcode: HU14 3HH

Contact Name: Martin

Test Date: 4 December 2018

Location: In Weighbridge

Manufacturer: VISHAY VT300

Serial/Plant No: T12310226

Capacity: 50000 kg

Increment: 20 kg

Frequency: 12 months

Next due: December 2019

As Found

Measurement Points	Reading	Deviation
0	0	0
8000	8000	0
16000	16000	0
23200	23220	20
33200	33220	20
43860	43900	40


Repeatability

43860	43860	0
-------	-------	---

Off-centre Loading

10000	10000	0
-------	-------	---

Engineer Name: J Scott



Definitive

Measurement Points	Reading	Deviation
0	0	0
8000	8000	0
16000	16000	0
23200	23200	0
33200	33200	0
43860	43860	0

Repeatability

43860	43860	0
-------	-------	---

Off-centre Loading

10000	10000	0
-------	-------	---

Trading Standards Maximum Tolerance

60	+/-	60	kg
----	-----	----	----

Traceable Weights Used: W18012

Calibration method and equipment used traceable to national standards.