



West End Grange  
 Epworth Water Gardens  
 West End Road  
 Epworth  
 Doncaster  
 DN9 1LA



Certificate Number:

2019552

Tel: 01427 875 867  
 Web: www.allweighs.co.uk

## Calibration Certificate of Accuracy

Business Name:

Address:

Postcode:

Contact Name:

Test Date:

Location:

Manufacturer:

Serial/Plant No:

Capacity:

Increment:

Frequency:

Next due:

### As Found

Measurement Points	Reading	Deviation
0	0	0
8000	8000	0
16000	16000	0
23940	23940	0
33940	33940	0
44780	44780	0

Repeatability

44780	44760	20
-------	-------	----

Off-centre Loading

10000	9960	40
-------	------	----

Engineer Name:

### Definitive

Measurement Points	Reading	Deviation
0	0	0
8000	8000	0
16000	16000	0
23940	23940	0
33940	33940	0
44780	44780	0

Repeatability

44780	44780	0
-------	-------	---

Off-centre Loading

10000	10000	0
-------	-------	---

Trading Standards Maximum Tolerance

60	+/-	60	kg
----	-----	----	----

Traceable Weights Used:

Calibration method and equipment used traceable to national standards.



West End Grange  
Epworth Water Gardens  
West End Road  
Epworth  
Doncaster  
DN9 1LA



Certificate Number:

2019553

Tel: 01427 875 867  
Web: www.allweighs.co.uk

## Calibration Certificate of Accuracy

Business Name: Transwaste

Address: Gibson Lane  
North Ferriby  
Melton

Postcode: HU14 3HH

Contact Name: Leanne Swann

Test Date: 10 December 2019

Location: Out Weighbridge

Manufacturer: Dini Argeo

Serial/Plant No: 0100676301

Capacity: 50000 kg

Increment: 20 kg

Frequency: 12 months

Next due: December 2020

### As Found

Measurement Points	Reading	Deviation
0	0	0
8000	8000	0
16000	16000	0
23940	23940	0
33940	33920	20
44780	44760	20

Repeatability

44780	44760	20
-------	-------	----

Off-centre Loading

10000	9980	20
-------	------	----

Engineer Name: J Scott

### Definitive

Measurement Points	Reading	Deviation
0	0	0
8000	8000	0
16000	16000	0
23940	23940	0
33940	33940	0
44780	44780	0

Repeatability

44780	44780	0
-------	-------	---

Off-centre Loading

10000	10000	0
-------	-------	---

Trading Standards Maximum Tolerance

60	+/-	60	kg
----	-----	----	----

Traceable Weights Used:

W19001

Calibration method and equipment used traceable to national standards.