

Allweighs Weighing Services
 Westend Grange
 West End Road
 Epworth
 Doncaster
 DN9 1LA
 Tel: 01427 875 867
 Web: www.allweighs.co.uk



Certificate No. KCS:

2021000065

Location: Weighbridge

Manufacturer: AWS/DINI ARGEO 3590

Serial/Plant No: 4515259

Capacity: 50000 kg

Increment: 20 kg

Frequency: 12 months

Test Date: 24 September 2021

Next due: September 2022

Calibration Certificate of Accuracy

Business Name: Transwaste

Address: Foster Street
Hull

Contact Name: HU8 8BT

Off-centre Loading:

| Load point: | Reading: | Deviation: | Load point: | Reading: | Deviation: |
|-------------|----------|------------|-------------|----------|------------|
| 1 | 11300 | 0 | 5 | 11300 | 0 |
| 2 | 11300 | 0 | 6 | 11300 | 0 |
| 3 | 11300 | 0 | | | |
| 4 | 11300 | 0 | | | |

Four loadcell
 Six loadcell

Single loadcell
 Eight loadcell

As Found

| Measurement Points: | Reading: | Deviation: |
|---------------------|----------|------------|
| 0 | 0 | 0 |
| 8000 | 8000 | 0 |
| 16000 | 16000 | 0 |
| 21240 | 21240 | 0 |
| 31240 | 31240 | 0 |
| 48580 | 48580 | 0 |

Repeatability

| | | |
|-------|-------|---|
| 48580 | 48580 | 0 |
|-------|-------|---|

Notes:

Engineer Name:

KS

Definitive

| Measurement Points: | Reading: | Deviation: |
|---------------------|----------|------------|
| 0 | 0 | 0 |
| 8000 | 8000 | 0 |
| 16000 | 16000 | 0 |
| 21240 | 21240 | 0 |
| 31240 | 31240 | 0 |
| 48580 | 48580 | 0 |

Repeatability

| | | |
|-------|-------|---|
| 48580 | 48580 | 0 |
|-------|-------|---|

Traceable Weights Used:

HTU

Trading Standards Maximum Tolerance:

60


+/-

60

kg

Calibration method and equipment used traceable to national standards.

QMF17 Issue 5 14/07/2020



Allweighs Weighing Services
 Westend Grange
 West End Road
 Epworth
 Doncaster
 DN9 1LA
 Tel: 01427 875 867
 Web: www.allweighs.co.uk



Certificate No. KCS:

2021000064

Location: Weighbridge

Manufacturer: AWS/DINI ARGEO DFWL

Serial/Plant No: 0100911902

Capacity: 50000 kg

Increment: 20 kg

Frequency: 12 months

Test Date: 24 September 2021

Next due: September 2022

Calibration Certificate of Accuracy

Business Name: Transwaste

Address: Melton Waste Park
 Gibson Lane
 Melton
 Hull

Contact Name: HU14 3HH

Off-centre Loading:

| Load point: | Reading: | Deviation: | Load point: | Reading: | Deviation: |
|-------------|----------|------------|-------------|----------|------------|
| 1 | 11300 | 0 | 5 | 11300 | 0 |
| 2 | 11300 | 0 | 6 | 11300 | 0 |
| 3 | 11300 | 0 | | | |
| 4 | 11300 | 0 | | | |

Four loadcell

Six loadcell

Single loadcell

Eight loadcell

As Found

| Measurement Points: | Reading: | Deviation: |
|---------------------|----------|------------|
| 0 | 0 | 0 |
| 8000 | 8000 | 0 |
| 16000 | 16000 | 0 |
| 21240 | 21240 | 0 |
| 31240 | 31240 | 0 |
| 48560 | 48560 | 0 |

Repeatability

| | | |
|-------|-------|---|
| 48560 | 48560 | 0 |
|-------|-------|---|

Notes:

Engineer Name:

KS

Definitive

| Measurement Points: | Reading: | Deviation: |
|---------------------|----------|------------|
| 0 | 0 | 0 |
| 8000 | 8000 | 0 |
| 16000 | 16000 | 0 |
| 21240 | 21240 | 0 |
| 31240 | 31240 | 0 |
| 48560 | 48560 | 0 |

Repeatability

| | | |
|-------|-------|---|
| 48560 | 48560 | 0 |
|-------|-------|---|

Traceable Weights Used:

HTU

Trading Standards Maximum Tolerance:

60


+/-

60

kg

Calibration method and equipment used traceable to national standards.

QMF17 Issue 5 14/07/2020



Allweighs Weighing Services
 Westend Grange
 West End Road
 Epworth
 Doncaster
 DN9 1LA
 Tel: 01427 875 867
 Web: www.allweighs.co.uk



Certificate No. KCS:

2021000063

Calibration Certificate of Accuracy

Business Name:

Address:

Contact Name:

Location:

Manufacturer:

Serial/Plant No:

Capacity:

Increment:

Frequency:

Test Date:

Next due:

Off-centre Loading:

| Load point: | Reading: | Deviation: | Load point: | Reading: | Deviation: |
|-------------|----------|------------|-------------|----------|------------|
| 1 | 11300 | 0 | 5 | 11300 | 0 |
| 2 | 11300 | 0 | 6 | 11300 | 0 |
| 3 | 11300 | 0 | | | |
| 4 | 11300 | 0 | | | |

Four loadcell Six loadcell

Single loadcell Eight loadcell

As Found

| Measurement Points: | Reading: | Deviation: |
|---------------------|----------|------------|
| 0 | 0 | 0 |
| 8000 | 8000 | 0 |
| 16000 | 16000 | 0 |
| 21240 | 21240 | 0 |
| 31240 | 31240 | 0 |
| 48560 | 48560 | 0 |

Repeatability

| | | |
|-------|-------|---|
| 48560 | 48560 | 0 |
|-------|-------|---|

Notes:

Engineer Name:

KS

Definitive

| Measurement Points: | Reading: | Deviation: |
|---------------------|----------|------------|
| 0 | 0 | 0 |
| 8000 | 8000 | 0 |
| 16000 | 16000 | 0 |
| 21240 | 21240 | 0 |
| 31240 | 31240 | 0 |
| 48560 | 48560 | 0 |

Repeatability

| | | |
|-------|-------|---|
| 48560 | 48560 | 0 |
|-------|-------|---|

Traceable Weights Used:

HTU

Trading Standards Maximum Tolerance:

60

+/-

60

kg

Calibration method and equipment used traceable to national standards.

QMF17 Issue 5 14/07/2020