



Allweighs Weighing Services
 Westend Grange
 West End Road
 Epworth
 Doncaster
 DN9 1LA
 Tel: 01427 875 867
 Web: www.allweighs.co.uk



Certificate No: KCS:

2022000027

Location: Weighbridge

Manufacturer: AWS / Dini Argeo DFW in bridge

Serial/Plant No: 0100911902

Capacity: 50000 kg

Increment: 20 kg

Frequency: 12 months

Test Date: 12 August 2022

Next due: August 2023

Calibration Certificate of Accuracy

Business Name: Transwaste

Address: Melton Waste Park
 Gibson Lane
 Melton
 Hull
 HU14 3HH

Contact Name: Graham Wildridge

Off-centre Loading:

Load point:	Reading:	Deviation:	Load point:	Reading:	Deviation:
1	11260	0	5	11260	
2	11260	0	6	11260	
3	11260	0			
4	11260	0			

Four loadcell

Six loadcell

Single loadcell

Eight loadcell

As Found

Measurement Points:	Reading:	Deviation:
0	0	0
8000	8000	0
16000	16000	0
20000	20000	0
35220	35220	0
48440	48440	0

Repeatability

23220	23220	0
-------	-------	---

Notes: All ok

Engineer Name:

JS

Definitive

Measurement Points:	Reading:	Deviation:
0	0	0
8000	8000	0
16000	16000	0
20000	20000	0
35220	35220	0
48440	48440	0

Repeatability


23220	23220	0
-------	-------	---

Traceable Weights Used: HTU

Trading Standards Maximum Tolerance: 60 +/- 60 kg

Calibration method and equipment used traceable to national standards.

QMF17 Issue 5 14/07/2020



Allweighs Weighing Services
 Westend Grange
 West End Road
 Epworth
 Doncaster
 DN9 1LA
 Tel: 01427 875 867
 Web: www.allweighs.co.uk



Certificate No: KCS:

2022000028

Location: Weighbridge

Manufacturer: AWS / Dini Argeo DFW out bridge

Serial/Plant No: 0702402559

Capacity: 50000 kg

Increment: 20 kg

Frequency: 12 months

Test Date: 12 August 2022

Next due: August 2023

Calibration Certificate of Accuracy

Business Name: Transwaste

Address: Melton Waste Park
 Gibson Lane
 Melton
 Hull
 HU14 3HH

Contact Name: Graham Wildridge

Off-centre Loading:

Load point:	Reading:	Deviation:	Load point:	Reading:	Deviation:
1	11260	0	5	11260	
2	11260	0	6	11260	
3	11260	0			
4	11260	0			

Four loadcell

Six loadcell

Single loadcell

Eight loadcell

As Found

Measurement Points:	Reading:	Deviation:
0	0	0
8000	8000	0
16000	16000	0
20000	20000	0
35220	35220	0
48440	48440	0

Repeatability

23220	23220	0
-------	-------	---

Notes: All ok

Engineer Name:

JS

Definitive

Measurement Points:	Reading:	Deviation:
0	0	0
8000	8000	0
16000	16000	0
20000	20000	0
35220	35220	0
48440	48440	0

Repeatability

23220	23220	0
-------	-------	---

Traceable Weights Used: HTU

Trading Standards Maximum Tolerance: 60 +/- 60 kg

Calibration method and equipment used traceable to national standards.

QMF17 Issue 5 14/07/2020