

Allweighs Weighing Services  
Westend Grange  
West End Road  
Epworth  
Doncaster  
DN9 1LA  
Tel: 01427 875 867  
Web: www.allweighs.co.uk



Certificate No. JS:

2023497

Location:

Out Weighbridge

Manufacturer:

Dini Argeo

Serial/Plant No:

0100911602

Capacity:

50000

kg

Increment:

20

kg

Frequency:

12

months

Test Date:

24 August 2023

Next due:

August 2024

## Calibration Certificate of Accuracy

Business Name:

Transwaste

Address:

Gibson Lane  
Melton Industrial Estate  
North Ferriby  
Hull  
HU14 3HH

Contact Name:



Graham Wildridge

### Off-centre Loading:

Load point:	Reading:	Deviation:	Load point:	Reading:	Deviation:
1	6680	0	5	6680	0
2	6680	0	6	6680	0
3	6680	0			
4	6680	0			



☐ Four loadcell

☒ Six loadcell



☐ Single loadcell

☐ Eight loadcell



### As Found

Measurement Points:	Reading:	Deviation:
0	0	0
8000	8000	0
16000	16000	0
23260	23260	0
33260	33260	0
43940	43960	20

Repeatability

43940	43940	0
-------	-------	---

Notes:

Weighbridge found to be within permitted tolerances.

### Definitive

Measurement Points:	Reading:	Deviation:
0	0	0
8000	8000	0
16000	16000	0
23260	23260	0
33260	33260	0
43940	43940	0

Repeatability

43940	43940	0
-------	-------	---

Traceable Weights Used:

W23017 + W23018

Trading  
Standards  
Maximum  
Tolerance:

60

+/-

60


kg

Calibration method and equipment used  
traceable to national standards.

QMF17 Issue 5 14/07/2020

Engineer  
Name:

J Scott



Allweighs Weighing Services  
Westend Grange  
West End Road  
Epworth  
Doncaster  
DN9 1LA  
Tel: 01427 875 867  
Web: www.allweighs.co.uk



Certificate No. JS:

2023498

Location:

In weighbridge

Manufacturer:

Dini Argeo

Serial/Plant No:

0702402559

Capacity:

50000

kg

Increment:

20

kg

Frequency:

12

months

Test Date:

24 August 2023

Next due:

August 2024

## Calibration Certificate of Accuracy

Business Name:

Transwaste

Address:

Gibson Lane  
Melton Industrial Estate  
North Ferriby  
Hull  
HU14 3HH

Contact Name:



Graham Wildridge

### Off-centre Loading:

Load point:	Reading:	Deviation:	Load point:	Reading:	Deviation:
1	6680	0	5	6680	0
2	6680	0	6	6660	20
3	6660	20			
4	6680	0			



☐ Four loadcell

☒ Six loadcell



☐ Single loadcell

☐ Eight loadcell



### As Found

Measurement Points:	Reading:	Deviation:
0	0	0
8000	8000	0
16000	16000	0
23260	23260	0
33260	33240	20
43940	43920	20

Repeatability

43940	43940	0
-------	-------	---

Notes:

Weighbridge found to be within tolerance.

### Definitive

Measurement Points:	Reading:	Deviation:
0	0	0
8000	8000	0
16000	16000	0
23260	23260	0
33260	33260	0
43940	43940	0

Repeatability

43940	43940	0
-------	-------	---

Traceable Weights Used:

W23017 + W23018

Trading  
Standards  
Maximum  
Tolerance:

60

+/-

60

kg

Calibration method and equipment used  
traceable to national standards.

QMF17 Issue 5 14/07/2020

Engineer  
Name:

J Scott