



Allweighs Weighing Services Ltd
Westend Grange
West End Road
Epworth
Doncaster
DN9 1LA
Carried out in partnership with
Shropshire Scale Company.
ISO: 9001-2015
Certificate No: GB25/000000026



Tel: 01427 875 867
Web: www.allweighs.co.uk

CJS

2026000237

Location: Weighbridge Inbound

Manufacturer: Dini Argeo DFWL-1

Serial/Plant No: 0100911602

Capacity: 50000 kg

Increment: 20 kg

Frequency: 12 months

Test Date: 23 June 2026

Next due: June 2027

Calibration Certificate of Accuracy

Business Name: Transwaste

Address: Melton Industrial Estate
Gibson Lane South
Melton
Hull
HU14 3HH

Contact Name: Graham Wildridge

Off-centre Loading Definitive: 6900

Load point:	Reading:	Deviation:	Load point:	Reading:	Deviation:
1	6900	0	5	6900	0
2	6900	0	6	6900	0
3	6900	0			
4	6900	0			

☐ Four loadcell	☒ Six loadcell														
<table><tr><td></td><td></td><td></td></tr><tr><td></td><td></td><td></td></tr></table>							<table><tr><td>2</td><td>4</td><td>6</td></tr><tr><td>1</td><td>3</td><td>5</td></tr></table>	2	4	6	1	3	5		
2	4	6													
1	3	5													
☐ Single loadcell	☐ Eight loadcell														
<table><tr><td></td><td></td><td></td></tr><tr><td></td><td></td><td></td></tr></table>							<table><tr><td></td><td></td><td></td><td></td></tr><tr><td></td><td></td><td></td><td></td></tr></table>								

As Found

Measurement Points:	Reading:	Deviation:
0	0	0
8000	8000	0
16000	16000	0
23920	23920	0
33920	33920	0
43540	43540	0

Repeatability

43540	43540	0
-------	-------	---

Notes: Weighbridge found to be within permitted tolerances.

Engineer Name:

CJS

C. Scott

Definitive

Measurement Points:	Reading:	Deviation:
0	0	0
8000	8000	0
16000	16000	0
23920	23920	0
33920	33920	0
43540	43540	0

Repeatability

43540	43540	0
-------	-------	---

Traceable Weights Used: TR13440

Trading Standards Maximum Tolerance:

60

+/-

60

kg

Calibration method and equipment used traceable to national standards.

AWSL:CC Issue 1 01/01/2025



Allweighs Weighing Services Ltd
Westend Grange
West End Road
Epworth
Doncaster
DN9 1LA
Carried out in partnership with
Shropshire Scale Company.
ISO: 9001-2015
Certificate No: GB25/000000026



Tel: 01427 875 867
Web: www.allweighs.co.uk

CJS

2026000236

Location: Weighbridge Outbound

Manufacturer: Dini Argeo DFWL-1

Serial/Plant No: 0702402559

Capacity: 50000 kg

Increment: 20 kg

Frequency: 12 months

Test Date: 23 June 2026

Next due: June 2027

Calibration Certificate of Accuracy

Business Name: Transwaste

Address: Melton Industrial Estate
Gibson Lane South
Melton
Hull
HU14 3HH

Contact Name: Graham Wildridge

Off-centre Loading Definitive: 6900

Load point:	Reading:	Deviation:	Load point:	Reading:	Deviation:
1	6900	0	5	6900	0
2	6900	0	6	6900	0
3	6900	0			
4	6900	0			

☐ Four loadcell	☒ Six loadcell
☐ Single loadcell	☐ Eight loadcell

As Found

Measurement Points:	Reading:	Deviation:
0	0	0
8000	8000	0
16000	16000	0
23920	23920	0
33920	33920	0
43540	43520	20

Repeatability

43540	43520	20
-------	-------	----

Definitive

Measurement Points:	Reading:	Deviation:
0	0	0
8000	8000	0
16000	16000	0
23920	23920	0
33920	33920	0
43540	43540	0

Repeatability

43540	43540	0
-------	-------	---

Notes: Weighbridge found to be within permitted tolerances.

Engineer Name:

CJS

C. Scott

Traceable Weights Used: TR13440

Trading Standards Maximum Tolerance:

60

+/-

60

kg

Calibration method and equipment used traceable to national standards.

AWSL:CC Issue 1 01/01/2025